

www.neb.com info@neb.com



## New England Biolabs Certificate of Analysis

Product Name: NEBuffer™ r3.1

Catalog Number: B6003S

Concentration: 10 X Concentrate

Packaging Lot Number: 10236778
Expiration Date: 02/2027
Storage Temperature: -20°C

Specification Version: PS-B6003S v1.0

Composition (1X): 50 mM Tris-HCl, 100 mM NaCl, 10 mM MgCl2, 100 μg/ml rAlbumin, (pH

7.9 @ 25°C)

NEBuffer™ r3.1 Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
B6003SVIAL	NEBuffer™ r3.1	10227733	Pass	

Assay Name/Specification	Lot # 10236778
Conductivity (buffers/solutions)	Pass
The conductivity of 10X NEBuffer™ r3.1 is between 84 and 101 mS/cm at 25°C.	
Endonuclease Activity (Nicking, Buffer) A 50 µl reaction in 1X NEBuffer™ r3.1 containing 1 µg of supercoiled PhiX174 DNA incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Functional Testing (Restriction Digest, Buffer) A 50 µl reaction in 1X NEBuffer™ r3.1 containing 1 µg of Lambda DNA and 1 unit of Asel incubated for 1 hour at 37°C results in complete digestion of the substrate DNA as determined by agarose gel electrophoresis.	Pass
Functional Testing (Restriction Digest, Buffer) A 50 μl reaction in 1X NEBuffer™ r3.1 containing 1 μg of pBC4 DNA and 1 unit of Notl incubated for 1 hour at 37°C results in complete digestion of the substrate DNA as determined by agarose gel electrophoresis.	Pass
Non-Specific DNase Activity (16 hour, Buffer) A 50 µl reaction in 1X NEBuffer™ r3.1 containing 1 µg of PhiX174-HaeIII DNA incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass



B6003S / Lot: 10236778

Page 1 of 2



Assay Name/Specification	Lot # 10236778
RNase Activity (Buffer) A 10 µl reaction in 1X NEBuffer™ r3.1 containing 40 ng of a 300 base single-stranded RNA is incubated at 37°C. After incubation for 16 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.	Pass
pH (buffers/solutions) The pH of 10X NEBuffer™ r3.1 is between pH 7.8 and 8.0 at 25°C.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Scott Schallehn
Production Scientist

I have

16 Feb 2024

Michael Tonello

Packaging Quality Control Inspector

25 Apr 2024



B6003S / Lot: 10236778

Page 2 of 2