

New England Biolabs Certificate of Analysis

Product Name: *Lambda PFG Ladder*
Catalog #: *N0341S*
Concentration: *50 µg/ml*
Unit Definition: *N/A*
Lot #: *0021606*
Assay Date: *06/2016*
Expiration Date: *06/2018*
Storage Temp: *-20°C*
Storage Conditions: *10 mM Tris-HCl, 1 mM EDTA, 50 % Glycerol, 1 % LMP Agarose, (pH 8.0 @ 25°C)*
Specification Version: *PS-N0341S v1.0*
Effective Date: *29 Jun 2016*

Assay Name/Specification (minimum release criteria)	Lot #0021606
Electrophoretic Pattern (Marker) - The banding pattern of Lambda PFG Ladder on a 1% CHEF PFGE gel shows clearly identifiable bands at each band of the marker, when stained with Ethidium Bromide at a concentration of 2.5 µg/ml.	Pass



Authorized by
Derek Robinson
29 Jun 2016



Inspected by
Vanessa Mathieu-Sheltry
09 Aug 2016

