

New England Biolabs Certificate of Analysis

Product Name: Authenticase™
Catalog Number: M0689S
Packaging Lot Number: 10206684
Expiration Date: 08/2025
Storage Temperature: -20°C
Storage Conditions: 500 mM NaCl, 10 mM Tris-HCl, 1 mM DTT, 0.1 mM EDTA, 50 % Glycerol, (pH 7.4 @ 25°C)
Specification Version: PS-M0689S/L v1.0

Authenticase™ Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
M0689SVIAL	Authenticase™	10196475	Pass
B2832SVIAL	Authenticase™ Reaction Buffer	10197704	Pass
B2831SVIAL	Authenticase™ DNA Annealing Buffer	10203840	Pass

Assay Name/Specification	Lot # 10206684
Functional Testing (Authenticase) A 20 µl reaction in 1X Authenticase Reaction Buffer containing 0.33 pmol of 60-mer heteroduplex of A/C, T/G, and 2 bp InDel mismatches, respectively, and a minimum of 1 µl of Authenticase is incubated at 42°C. After incubation for 30 minutes, >90% of the substrate DNA was digested as determined by agarose gel electrophoresis.	Pass
RNase Activity (Extended Digestion) A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA and a minimum of 1 µl of Authenticase is incubated at 37°C. After incubation for 2 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Jamie Souza
Production Scientist
24 Aug 2023



Josh Hersey
Packaging Quality Control Inspector
31 Aug 2023