

New England Biolabs Certificate of Analysis

Product Name: Q5® High-Fidelity PCR Kit
Catalog Number: E0555L
Packaging Lot Number: 10099477
Expiration Date: 09/2022
Storage Temperature: -20°C
Specification Version: PS-E0555S/L v3.0

Q5® High-Fidelity PCR Kit Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
N0550AAVIAL	Quick-Load® Purple 1 kb Plus DNA Ladder	10096449	Pass
M0492SVIAL	Q5® High-Fidelity 2X Master Mix	10083995	Pass
B7024SVIAL	Gel Loading Dye, Purple (6X)	10084974	Pass
B1502AVIAL	Nuclease-free Water	10096705	Pass

Assay Name/Specification	Lot # 10099477
<p>* Individual Product Component Note Standard Quality Control Tests are performed for each component included in Q5® High-Fidelity PCR Kit and meet the designated specifications.</p>	Pass
<p>PCR Amplification (20 kb Lambda DNA, Master Mix) A 50 µl reaction in 1X Q5® High-Fidelity Master Mix and 1.0 µM primers containing 10 ng Lambda DNA for 22 cycles of PCR amplification results in the expected 20 kb product.</p>	Pass
<p>PCR Amplification (7 kb Human Genomic DNA, Master Mix) A 50 µl reaction in 1X Q5® High-Fidelity Master Mix and 0.5 µM primers containing 20 ng Human Genomic DNA for 30 cycles of PCR amplification results in the expected 7 kb product.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Christie Vazquez

Christie Vazquez
Production Scientist
09 Feb 2021

Josh Hersey

Josh Hersey
Packaging Quality Control Inspector
09 Feb 2021