

New England Biolabs Certificate of Analysis


Product Name: OneTaq® One-Step RT-PCR Kit
 Catalog Number: E5315S
 Lot Number: 10054117
 Expiration Date: 08/2020
 Storage Temperature: -20°C
 Specification Version: PS-E5315S v1.0

OneTaq® One-Step RT-PCR Kit Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
M1205AVIAL	Quick Load® OneTaq® One-Step Reaction Mix	10049434	Pass
M1204AVIAL	OneTaq® One-Step Reaction Mix	10049433	Pass
M1203AVIAL	OneTaq® One-Step Enzyme Mix	10049432	Pass
B1502AVIAL	Nuclease-free Water	10046857	Pass

Assay Name/Specification	Lot # 10054117
<p>* Individual Product Component Note Standard Quality Control Tests are performed for each component included in OneTaq® One-Step RT-PCR Kit and meet the designated specifications.</p>	Pass
<p>Functional Testing (RT-PCR, Sensitivity) A 50 µl reaction in OneTaq® One-Step Reaction Mix or Quick-Load® OneTaq® One-Step Reaction Mix containing 10 pg of total RNA, 1X OneTaq® One-Step Enzyme Mix, and 0.4 µM primers for 40 cycles of PCR amplification results in the expected 566 bp product.</p>	Pass
<p>Functional Testing (RT-PCR, Multiplexing) A 50 µl reaction in OneTaq® One-Step Reaction Mix or Quick-Load® OneTaq® One-Step Reaction Mix containing 10 ng of total RNA, 1X OneTaq® One-Step Enzyme Mix, and 0.1 µM primers for 40 cycles of PCR amplification results in the expected three products.</p>	Pass
<p>Functional Testing (RT-PCR, Length) A 50 µl reaction in OneTaq® One-Step Reaction Mix or Quick-Load® OneTaq® One-Step Reaction Mix containing 100 ng of total RNA, 1X OneTaq® One-Step Enzyme Mix, and 0.4 µM primers for 40 cycles of PCR amplification results in the expected 9.3 kb product.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.





Doreen Duquette
Production Scientist
16 Aug 2019



Michael Tonello
Packaging Quality Control Inspector
04 Oct 2019