

New England Biolabs Certificate of Analysis

Product Name: NEBExpress[®] Cell-Free *E. coli* Protein Synthesis System
 Catalog Number: E5360L
 Packaging Lot Number: 10129585
 Expiration Date: 11/2023
 Storage Temperature: -80°C
 Specification Version: PS-E5360S/L v2.0


NEBExpress [®] Cell-Free <i>E. coli</i> Protein Synthesis System Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
P0864SVIAL	NEBExpress [®] S30 Synthesis Extract	10129525	Pass
N3273AVIAL	NEBExpress [®] Control DHFR-His Plasmid	10103266	Pass
M1019AAVIAL	T7 RNA Polymerase	10129547	Pass
M1018AAVIAL	RNase Inhibitor, Murine	10129546	Pass
B0864SVIAL	Protein Synthesis Buffer (2X)	10129524	Pass

Assay Name/Specification	Lot # 10129585
<p>* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBExpress[®] Cell-Free <i>E. coli</i> Protein Synthesis System and meet the designated specifications.</p>	Pass
<p>Functional Testing (Cell-Free Protein Synthesis Assay) (vGFP, Fluorescence) A 50 µl reaction in the presence of 250 ng <i>E. coli</i> vGFP template DNA containing the components of the NEBExpress[®] Cell-Free <i>E. coli</i> Protein Synthesis System incubated for 5 hours at 37°C results in the expected product as determined by a fluorescence signal ≥ 2 times higher than the reference standard signal.</p>	Pass
<p>Functional Testing (Cell-Free Protein Synthesis Assay) (vGFP, SDS-PAGE) A 50 µl reaction in the presence of 250 ng <i>E. coli</i> vGFP template DNA containing the components of the NEBExpress[®] Cell-Free <i>E. coli</i> Protein Synthesis System incubated for 3 hours at 37°C results in the expected 28 kDa product as determined by SDS-PAGE with Coomassie Blue detection.</p>	Pass
<p>Functional Testing (Cell-Free Protein Synthesis Assay) (DHFR) A 50 µl reaction in the presence of 250 ng NEBExpress[®] Control DHFR-His Plasmid containing the components of the NEBExpress[®] Cell-Free <i>E. coli</i> Protein Synthesis</p>	Pass


Assay Name/Specification	Lot # 10129585
System incubated for 3 hours at 37°C results in the expected 20 kDa product as determined by SDS-PAGE with Coomassie Blue detection.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Cory Tuckey
Production Scientist
17 Nov 2021



Michael Tonello
Packaging Quality Control Inspector
17 Nov 2021