

## New England Biolabs Certificate of Analysis

Product Name: NEBNext® Ultra™ Directional RNA Library Prep Kit for Illumina®  
 Catalog Number: E7420L  
 Lot Number: 10026026  
 Expiration Date: 07/2019  
 Storage Temperature: -20°C  
 Specification Version: PS-E7420S/L v1.0

NEBNext® Ultra™ Directional RNA Library Prep Kit for Illumina® Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7431AAVIAL	Nuclease-free Water	0011801	Pass
E7428AAVIAL	NEBNext® USER® Enzyme	0161801	Pass
E7426AAVIAL	NEBNext® Second Strand Synthesis Reaction Buffer (dUTP Mix)	0121801	Pass
E7425AAVIAL	NEBNext® Second Strand Synthesis Enzyme Mix	0161801	Pass
E7424AAVIAL	RNase Inhibitor, Murine	0311801	Pass
E7423AAVIAL	ProtoScript® II Reverse Transcriptase	0041801	Pass
E7422AAVIAL	Random Primers	0211801	Pass
E7421AAVIAL	NEBNext® First Strand Synthesis Reaction Buffer	0151801	Pass
E7373AAVIAL	Blunt/TA Ligase Master Mix	0371801	Pass
E7372AAVIAL	NEBNext® End Repair Reaction Buffer	0211801	Pass
E7371AAVIAL	NEBNext® End Prep Enzyme Mix	0151801	Pass
E6625AAVIAL	NEBNext® Q5® Hot Start HiFi PCR Master Mix	0081801	Pass

Assay Name/Specification	Lot # 10026026
<p><b>Functional Testing (Library Construction, RNA)</b>            Each set of reagents is functionally validated and compared to a previous lot through construction of libraries made from commercially available RNA, using the kit's minimum and maximum input requirements. Libraries made from previous and current lots for both input RNA amounts are sequenced together on the same Illumina® flow cell and compared across various metrics including library yield, individual transcript abundance correlations (low vs. high input, old lot vs. new lot), 5'-3' transcript coverage, and fraction of reads mapping to a reference.</p> <p><b>* Individual Product Component Note</b>            Standard Quality Control Tests are performed for each component included in NEBNext® Ultra™ Directional RNA Library Prep Kit for Illumina® and meet the designated</p>	<p>Pass</p> <p>Pass</p>

Assay Name/Specification	Lot # 10026026
specifications.	

This product has been tested and shown to be in compliance with all specifications.



---

Christine Sumner  
Production Scientist  
18 Oct 2018



---

Michael Tonello  
Packaging Quality Control Inspector  
18 Oct 2018