

New England Biolabs Certificate of Analysis

Product Name: NEBNext® Ultra™ Directional RNA Library Prep Kit for Illumina®
 Catalog Number: E7420S
 Lot Number: 10036065
 Expiration Date: 02/2020
 Storage Temperature: -20°C
 Specification Version: PS-E7420S/L v1.0

NEBNext® Ultra™ Directional RNA Library Prep Kit for Illumina® Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7431AVIAL	Nuclease-free Water	10019144	Pass
E7428AVIAL	NEBNext® USER® Enzyme	10019145	Pass
E7426AVIAL	NEBNext® Second Strand Synthesis Reaction Buffer (dUTP Mix)	10019140	Pass
E7425AVIAL	NEBNext® Second Strand Synthesis Enzyme Mix	10019139	Pass
E7424AVIAL	RNase Inhibitor, Murine	10019138	Pass
E7423AVIAL	ProtoScript® II Reverse Transcriptase	10019146	Pass
E7422AVIAL	Random Primers	10019137	Pass
E7421AVIAL	NEBNext® First Strand Synthesis Reaction Buffer	10019136	Pass
E7373AVIAL	Blunt/TA Ligase Master Mix	10019143	Pass
E7372AVIAL	NEBNext® End Repair Reaction Buffer	10019142	Pass
E7371AVIAL	NEBNext® End Prep Enzyme Mix	10019141	Pass
E6625AVIAL	NEBNext® Q5® Hot Start HiFi PCR Master Mix	10019147	Pass

Assay Name/Specification	Lot # 10036065
<p>* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext® Ultra™ Directional RNA Library Prep Kit for Illumina® and meet the designated specifications.</p>	Pass
<p>Functional Testing (Library Construction, RNA) Each set of reagents is functionally validated and compared to a previous lot through construction of libraries made from commercially available RNA, using the kit's minimum and maximum input requirements. Libraries made from previous and current lots for both input RNA amounts are sequenced together on the same Illumina® flow cell and compared across various metrics including library yield, individual transcript abundance correlations (low vs. high input, old lot vs. new lot), 5'-3'</p>	Pass

Assay Name/Specification	Lot # 10036065
transcript coverage, and fraction of reads mapping to a reference.	

This product has been tested and shown to be in compliance with all specifications.



Christine Sumner
Production Scientist
07 Feb 2019



Michael Tonello
Packaging Quality Control Inspector
07 Feb 2019