

New England Biolabs Certificate of Analysis

Product Name: NEBNext[®] Ultra[™] Directional RNA Library Prep Kit for Illumina[®]
 Catalog Number: E7420S
 Packaging Lot Number: 10078595
 Expiration Date: 08/2021
 Storage Temperature: -20°C
 Specification Version: PS-E7420S/L v1.0

NEBNext [®] Ultra [™] Directional RNA Library Prep Kit for Illumina [®] Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7431AVIAL	Nuclease-free Water	10063939	Pass
E7428AVIAL	NEBNext [®] USER [®] Enzyme	10063940	Pass
E7426AVIAL	NEBNext [®] Second Strand Synthesis Reaction Buffer (dUTP Mix)	10063935	Pass
E7425AVIAL	NEBNext [®] Second Strand Synthesis Enzyme Mix	10063934	Pass
E7424AVIAL	RNase Inhibitor, Murine	10063933	Pass
E7423AVIAL	ProtoScript [®] II Reverse Transcriptase	10063942	Pass
E7422AVIAL	Random Primers	10063932	Pass
E7421AVIAL	NEBNext [®] First Strand Synthesis Reaction Buffer	10063931	Pass
E7373AVIAL	Blunt/TA Ligase Master Mix	10063938	Pass
E7372AVIAL	NEBNext [®] End Repair Reaction Buffer	10063937	Pass
E7371AVIAL	NEBNext [®] End Prep Enzyme Mix	10063936	Pass
E6625AVIAL	NEBNext [®] Q5 [®] Hot Start HiFi PCR Master Mix	10063943	Pass

Assay Name/Specification	Lot # 10078595
<p>Functional Testing (Library Construction, RNA) Each set of reagents is functionally validated and compared to a previous lot through construction of libraries made from commercially available RNA, using the kit's minimum and maximum input requirements. Libraries made from previous and current lots for both input RNA amounts are sequenced together on the same Illumina[®] flow cell and compared across various metrics including library yield, individual transcript abundance correlations (low vs. high input, old lot vs. new lot), 5'-3' transcript coverage, and fraction of reads mapping to a reference.</p> <p>* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext[®] Ultra[™] Directional RNA Library Prep Kit for Illumina[®] and meet the designated</p>	<p>Pass</p> <p>Pass</p>

Assay Name/Specification	Lot # 10078595
specifications.	

This product has been tested and shown to be in compliance with all specifications.



Christine Sumner
Production Scientist
30 Jun 2020



Michael Tonello
Packaging Quality Control Inspector
30 Jun 2020