

New England Biolabs Certificate of Analysis

Product Name: NEBNext® Ultra™ II End Repair/dA-Tailing Module
Catalog Number: E7546L
Lot Number: 10048654
Expiration Date: 03/2020
Storage Temperature: -20°C
Specification Version: PS-E7546S/L v1.0

NEBNext® Ultra™ II End Repair/dA-Tailing Module Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7647AAVIAL	NEBNext® Ultra™ II End Prep Reaction Buffer	10040648	Pass
E7646AAVIAL	NEBNext® Ultra™ II End Prep Enzyme Mix	10040647	Pass

Assay Name/Specification	Lot # 10048654
Functional Testing (Library Construction, DNA Modules) Each set of reagents is functionally validated and compared to a previous lot through construction of DNA libraries. Libraries made from previous and current lots for the recommended input DNA amounts are sequenced together on the same Illumina® flow cell and compared across various metrics including library yield, fraction of reads aligning to the reference, GC bias, and insert size.	Pass
* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext® Ultra™ II End Repair/dA-Tailing Module and meet the designated specifications.	Pass

This product has been tested and shown to be in compliance with all specifications.



Christine Sumner
Production Scientist
25 Jun 2019



Michael Tonello
Packaging Quality Control Inspector
25 Jun 2019