

## New England Biolabs Certificate of Analysis

**Product Name:** Amylose Resin High Flow  
**Catalog Number:** E8022S  
**Packaging Lot Number:** 10065126  
**Expiration Date:** 08/2022  
**Storage Temperature:** 4°C  
**Specification Version:** PS-E8022S/L v1.0

Amylose Resin High Flow Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E8022SVIAL	Amylose Resin High Flow	10053612	Pass

Assay Name/Specification	Lot # 10065126
<p><b>Functional Binding Assay (Resin Binding Capacity)</b>            Amylose Resin High Flow ( 1 ml ) was packed into a column and equilibrated with column buffer. Crude extract from E. coli containing a plasmid that expresses a MBP5*-paramyosinΔSal fusion protein ( 8 ml ) was then passed through the column at 25°C, then washed with column buffer and the target protein eluted with 4 ml of column buffer containing 10 mM maltose. Binding capacity was determined to be &gt;4 mg MBP5*-paramyosinΔSal /ml of resin based on A280 of the eluate.</p>	<b>Pass</b>
<p><b>Functional Binding Assay (Resin Binding Specificity)</b>            Amylose Resin High Flow ( 1 ml ) was packed into a column and equilibrated with column buffer. Crude extract from E. coli containing a plasmid that expresses a MBP5*-paramyosinΔSal fusion protein ( 8 ml ) was then passed through the column at 25°C, and then washed with column buffer. The target protein was eluted with 4 ml of column buffer containing 10 mM maltose. SDS-PAGE of the eluate on a 10-20% Tris-Glycine gel confirms low non-specific binding of extract proteins and high isolation of target.</p>	<b>Pass</b>

This product has been tested and shown to be in compliance with all specifications.



---

Brad Landgraf  
Production Scientist  
30 Aug 2019



---

Michael Tonello  
Packaging Quality Control Inspector  
30 Jan 2020