

New England Biolabs Certificate of Analysis

Product Name: *Ph.D.™-12 Phage Display Peptide Library Kit*
 Catalog Number: *E8110S*
 Packaging Lot Number: *10165302*
 Expiration Date: *09/2023*
 Storage Temperature: *-20°C*
 Specification Version: *PS-E8110S v2.0*

Ph.D.™-12 Phage Display Peptide Library Kit Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
S1259AVIAL	-96 gIII Sequencing Primer (20-mer)	10159998	Pass
S1258AVIAL	-28 gIII Sequencing Primer (22-mer)	10126854	Pass
N7024AVIAL	Biotin	10111664	Pass
N7023AVIAL	Streptavidin, lyophilized	10115978	Pass
E8111AVIAL	Ph.D.™-12 Phage Display Peptide Library	10111203	Pass
E4104SVIAL	E.coli K12 ER2738	10124262	Pass

Assay Name/Specification	Lot # 10165302
Absolute Phage Titer Infection of a mid-log culture of E. coli ER2738 with Ph.D.™-12 Phage Display Peptide Library followed by plating, yields $\geq 1 \times 10^{13}$ pfu/ml.	Pass
Functional Testing (Panning) A 100-fold representation of the Ph.D.™-12 Phage Display Peptide Library containing approximately 10 ¹¹ pfu is diluted in 200 μ l TBS and panned against 300 ng β -endorphin monoclonal antibody. The bound phage is affinity captured using magnetic beads and eluted with 1 ml of 0.2M Glycine-HCl, pH 2.2. After three rounds of selection, $\geq 75\%$ of sequences contain a motif related to the known epitope for the antibody.	Pass
Sequence Verification (DNA) The Ph.D.™-12 Phage Display Peptide Library was sequenced using 5'-CCCATGTACCGTAACACTGAGTTTC-3' as a primer to confirm the correct form of the cloned insert on the displayed peptide, X12-GGG.	Pass
Phage Contamination (Environmental) A 1:100 dilution of an overnight culture of E. coli ER2738 was made in 20 ml LB, to which 10 ³ pfu of Ph.D.™-12 Phage Display Peptide Library was added. The flask was	Pass

Assay Name/Specification	Lot # 10165302
incubated at 37°C on a rotating shaker for 5 hours. A 1 ml volume of culture was removed and centrifuged. A volume of culture supernatant equivalent to the initial PFU input was added to a second, 20 ml culture like the first. The final culture supernatant was plated on three LB/IPTG/Xgal plates and then titered. Fewer than 20% clear or white plaques were observed in a minimum of 100 total plaques counted on each plate.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Beth Paschal
Production Scientist
24 Sep 2021



Michael Tonello
Packaging Quality Control Inspector
15 Dec 2022