

## New England Biolabs Certificate of Analysis

**Product Name:** *Phusion® High-Fidelity PCR Master Mix with GC Buffer*  
**Catalog Number:** *M0532S*  
**Concentration:** *2 X Concentrate*  
**Packaging Lot Number:** *10065670*  
**Expiration Date:** *06/2021*  
**Storage Temperature:** *-20°C*  
**Specification Version:** *PS-M0532S/L v2.0*  
**Composition (1X):** *1 X Phusion® GC Buffer, 0.2 mM dATP, 0.2 mM dCTP, 0.2 mM dGTP, 0.2 mM dTTP, 20 units/ml Phusion® High-Fidelity DNA Polymerase*

Phusion® High-Fidelity PCR Master Mix with GC Buffer Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
M0532SVIAL	Phusion® High-Fidelity PCR Master Mix with GC Buffer	10062932	Pass
B0515AVIAL	DMSO	10062895	Pass

Assay Name/Specification	Lot # 10065670
<b>PCR Amplification (20 kb Lambda DNA, Master Mix)</b> A 50 µl reaction in 1X Phusion® High-Fidelity PCR Master Mix with GC Buffer and 1.0 µM primers containing 10 ng Lambda DNA for 22 cycles of PCR amplification results in the expected 20 kb product.	Pass
<b>PCR Amplification (7.5 kb Human Genomic DNA, Master Mix)</b> A 50 µl reaction in 1X Phusion® High-Fidelity PCR Master Mix with GC Buffer and 1.0 µM primers containing 50 ng Human Genomic DNA for 30 cycles of PCR amplification results in the expected 7.5 kb product.	Pass

This product has been tested and shown to be in compliance with all specifications.

*Christie Vazquez*

Christie Vazquez  
Production Scientist  
10 Jan 2020

*Josh Hersey*

Josh Hersey  
Packaging Quality Control Inspector  
17 Mar 2020