

## New England Biolabs Certificate of Analysis

**Product Name:** Luna<sup>®</sup> Probe One-Step RT-qPCR 4X Mix with UDG (No ROX)  
**Catalog Number:** M3029E  
**Concentration:** 4 X Concentrate  
**Packaging Lot Number:** 10148078  
**Expiration Date:** 04/2023  
**Storage Temperature:** -20°C  
**Specification Version:** PS-M3029S/L/E v1.0  
**Composition (1X):** Proprietary

Luna <sup>®</sup> Probe One-Step RT-qPCR 4X Mix with UDG (No ROX) Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
M3029EVIAL	Luna Probe One-Step RT-qPCR 4X Mix with UDG (No ROX)	10148079	Pass
B1502EVIAL	Nuclease-free Water	10137881	Pass

Assay Name/Specification	Lot # 10148078
<b>RNase Activity (Extended Digestion)</b> A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA and a minimum of 1 µl of Luna <sup>®</sup> Probe One-Step RT-qPCR 4X Mix with UDG (No ROX) is incubated at 37°C. After incubation for 4 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.	Pass
<b>qPCR DNA Contamination (E. coli Genomic)</b> A minimum of 1 µl of Luna <sup>®</sup> Probe One-Step RT-qPCR 4X Mix with UDG (No ROX) is screened for the presence of E. coli genomic DNA using SYBR <sup>®</sup> Green qPCR with primers specific for the E. coli 16S rRNA locus. Results are quantified using a standard curve generated from purified E. coli genomic DNA. The measured level of E. coli genomic DNA contamination is ≤ 1 E. coli genome.	Pass
<b>Non-Specific DNase Activity (16 hour, Buffer)</b> A 50 µl reaction in 1X Luna <sup>®</sup> Probe One-Step RT-qPCR Mix with UDG (No ROX) containing 1 µg of T3 or T7 DNA in addition to a reaction containing Lambda-HindIII DNA incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
<b>Functional Testing (One-Step RT-qPCR)</b>	Pass

Assay Name/Specification	Lot # 10148078
Luna® Probe One-Step RT-qPCR 4X Mix with UDG (No ROX) is functionally tested in one-step RT-qPCR with human RNA template, resulting in a standard curve with a calculated qPCR efficiency of 90-110%, and a dynamic range of 8 orders of magnitude.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.



Christie Vazquez  
Production Scientist  
17 May 2022



Josh Hersey  
Packaging Quality Control Inspector  
17 May 2022