

## New England Biolabs Certificate of Analysis

**Product Name:** Quick-Load® 1 kb DNA Ladder  
**Catalog Number:** N0468S  
**Concentration:** 50 µg/ml  
**Unit Definition:** N/A  
**Packaging Lot Number:** 10168736  
**Expiration Date:** 10/2024  
**Storage Temperature:** 4°C  
**Storage Conditions:** 3.3 mM Tris-HCl (pH 8.0), 11 mM EDTA , 0.015 % Bromophenol Blue , 0.017 % SDS , 2.5 % Ficoll 400  
**Specification Version:** PS-N0468S/L v1.0

| Quick-Load® 1 kb DNA Ladder Component List |                             |            |                      |
|--|-----------------------------|------------|----------------------|
| NEB Part Number                            | Component Description       | Lot Number | Individual QC Result |
| N0468SVIAL                                 | Quick-Load® 1 kb DNA Ladder | 10168734   | Pass                 |

| Assay Name/Specification   | Lot # 10168736 |
|--|----------------|
| <p><b>Non-Specific DNase Activity (DNA, 16 hour)</b><br/>           A 50 µl reaction in 1X NEBuffer 2 containing 2.5 µg of Quick-Load® 1 kb DNA Ladder incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.</p> | Pass           |
| <p><b>Electrophoretic Pattern (Marker)</b><br/>           The banding pattern of Quick-Load® 1 kb DNA Ladder on a 1.2% agarose gel shows discrete, clearly identifiable bands at each band of the marker, when stained with Ethidium Bromide at a concentration of 0.5 µg/ml.</p>                        | Pass           |
| <p><b>DNA Concentration (A260)</b><br/>           The concentration of Quick-Load® 1 kb DNA Ladder is between 50 and 55 µg/ml as determined by UV absorption at 260 nm.</p>  | Pass           |

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.



Vanessa Mathieu-Sheltry  
Production Scientist  
04 Nov 2022



Michael Tonello  
Packaging Quality Control Inspector  
08 Nov 2022