

New England Biolabs Certificate of Analysis

Product Name: *λ DNA-Mono Cut Mix*
Catalog #: *N3019S/L*
Concentration: *500 µg/ml*
Unit Definition: *N/A*
Lot #: *0191608*
Assay Date: *08/2016*
Expiration Date: *8/2018*
Storage Temp: *-20°C*
Storage Conditions: *10 mM Tris-HCl (pH 8.0), 1 mM EDTA*
Specification Version: *PS-N3019S/L v1.0*
Effective Date: *25 Jun 2014*

Assay Name/Specification (minimum release criteria)	Lot #0191608
A260/A280 Assay - The ratio of UV absorption of λ DNA-Mono Cut Mix at 260 and 280 nm is between 1.8 and 2.0.	Pass
DNA Concentration (A260) - The concentration of λ DNA-Mono Cut Mix is between 500 and 550 µg/ml as determined by UV absorption at 260 nm.	Pass
Electrophoretic Pattern (PFGE Marker) - The banding pattern of λ DNA-Mono Cut Mix on a 1% CHEF PFG gel shows discrete, clearly identifiable bands at each band of the marker, as determined by gel electrophoresis using Ethidium Bromide.	Pass
Non-Specific DNase Activity (DNA, 16 hour) - A 50 µl reaction in 1X NEBuffer 2 containing 2.5 µg of λ DNA-Mono Cut Mix incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass



Authorized by
Derek Robinson
25 Jun 2014



Inspected by
Vanessa Mathieu-Sheltry
23 Aug 2016

