

## New England Biolabs Certificate of Analysis

**Product Name:** *Color Prestained Protein Standard, Broad Range (11-245 kDa)*  
**Catalog Number:** *P7712S*  
**Concentration:** *0.2 mg/ml*  
**Lot Number:** *10028238*  
**Expiration Date:** *08/2020*  
**Storage Temperature:** *-20°C*  
**Storage Conditions:** *20 mM Tris-Phosphate, 2 % SDS, 0.2 mM DTT, 3.6 M Urea, 15 % Glycerol, (pH 7.5 @ 25°C)*  
**Specification Version:** *PS-P7712S/L v1.0*

Color Prestained Protein Standard, Broad Range (11-245 kDa) Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
P7712SVIAL	Color Prestained Protein Standard, Broad Range (11-245 kDa)	10016568	Pass

Assay Name/Specification	Lot # 10028238
<b>Electrophoretic Pattern (Protein Marker)</b> The banding pattern of Color Prestained Protein Standard, Broad Range (11-245 kDa) on a 10-20% Tris-Glycine gel shows discrete, clearly identifiable bands at each band of the marker, as determined by gel electrophoresis using Coomassie Blue detection.	Pass
<b>Protease Activity (Protein Markers &amp; Ladders)</b> A 20 µl sample of Color Prestained Protein Standard, Broad Range (11-245 kDa) incubated for 16 hours at 37°C, results in no detectable degradation of the protein mixture as determined by SDS-PAGE with Coomassie Blue detection.	Pass

This product has been tested and shown to be in compliance with all specifications.

*Beth M. Paschal*

---

Beth Paschal  
Production Scientist  
16 Aug 2018

*Michael Tonello*

---

Michael Tonello  
Packaging Quality Control Inspector  
05 Nov 2018