

New England Biolabs Certificate of Analysis

Product Name: Xmal
Catalog Number: R0180M
Concentration: 50,000 U/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg adenovirus-2 in 1 hour at 37°C in a total reaction volume of 50 µl.
Lot Number: 10021272
Expiration Date: 09/2020
Storage Temperature: -20°C
Storage Conditions: 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA
Specification Version: PS-R0180M v1.0

Xmal Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0180MVIAl	Xmal	10020873	Pass
B7204SVIAl	CutSmart® Buffer	10015393	Pass
B7024SVIAl	Gel Loading Dye, Purple (6X)	10007497	Pass

Assay Name/Specification	Lot # 10021272
<p>Endonuclease Activity (Nicking) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of supercoiled PhiX174 DNA and a minimum of 50 Units of Xmal incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.</p>	Pass
<p>Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 50 units of Xmal incubated for 4 hours at 37°C releases <0.2% of the total radioactivity.</p>	Pass
<p>Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of pXba DNA with Xmal, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with Xmal.</p>	Pass
<p>Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of pXba DNA and a minimum of 10 Units of Xmal incubated for 16 hours at 37°C results in a DNA pattern free of</p>	Pass

Assay Name/Specification	Lot # 10021272
<p>detectable nuclease degradation as determined by agarose gel electrophoresis.</p> <p>Protein Purity Assay (SDS-PAGE) Xmal is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.</p>	<p>Pass</p>

This product has been tested and shown to be in compliance with all specifications.



Tony Spear-Alfonso
Production Scientist
14 May 2018



Michael Tonello
Packaging Quality Control Inspector
24 Sep 2018