

New England Biolabs Certificate of Analysis

Product Name: NotI
Catalog Number: R0189M
Concentration: 50,000 U/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of pBC4 DNA in 1 hour at 37°C in a total reaction volume of 50 µl.
Packaging Lot Number: 10065830
Expiration Date: 02/2022
Storage Temperature: -20°C
Storage Conditions: 250 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 0.15% Triton X-100, 200 µg/ml BSA
Specification Version: PS-R0189M v1.0

NotI Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0189MVIAL	NotI	10065829	Pass
B7203SVIAL	NEBuffer™ 3.1	10053974	Pass
B7024SVIAL	Gel Loading Dye, Purple (6X)	10065745	Pass

Assay Name/Specification	Lot # 10065830
Endonuclease Activity (Nicking) A 50 µl reaction in NEBuffer 3.1 containing 1 µg of supercoiled PhiX174 DNA and a minimum of 100 Units of NotI incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in NEBuffer 3.1 containing 1 µg of a mixture of single and double-stranded [³ H] E. coli DNA and a minimum of 100 units of NotI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of pBC4 DNA with NotI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with NotI.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in NEBuffer 3.1 containing 1 µg of pBC4 DNA and a minimum of 100 Units of NotI incubated for 16 hours at 37°C results in a DNA pattern free of	Pass

Assay Name/Specification	Lot # 10065830
detectable nuclease degradation as determined by agarose gel electrophoresis.	

This product has been tested and shown to be in compliance with all specifications.



Stephanie Cornelio
Production Scientist
30 Jan 2020



Michael Tonello
Packaging Quality Control Inspector
19 Mar 2020