

New England Biolabs Certificate of Analysis


Product Name: *DrdI*
Catalog #: *R0530S/L*
Concentration: *5,000 units/ml*
Unit Definition: *One unit is defined as the amount of enzyme required to digest 1 µg of pUC19 DNA in 1 hour at 37°C in a total reaction volume of 50 µl.*
Lot #: *0191507*
Assay Date: *07/2015*
Expiration Date: *7/2017*
Storage Temp: *-20°C*
Storage Conditions: *50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA*
Specification Version: *PS-R0530S/L v2.0*
Effective Date: *02 Dec 2013*

| Assay Name/Specification (minimum release criteria) | Lot #0191507 |
|---|--------------|
| Exonuclease Activity (Radioactivity Release) - A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³ H] <i>E. coli</i> DNA and a minimum of 50 units of DrdI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity. | Pass |
| Ligation and Recutting (Terminal Integrity) - After a 10-fold over-digestion of pUC19 DNA with DrdI, ~75% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with DrdI. | Pass |
| Non-Specific DNase Activity (16 Hour) - A 50 µl reaction in CutSmart™ Buffer containing 1 µg of pUC19 DNA and a minimum of 25 Units of DrdI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis. | Pass |

* The BSA in this product has been granted an EDQM "Certificate of Suitability" from the European Directorate for the Quality of Medicines (# R1-CEP-2003-204-Rev00) and has been granted a USDA Certificate for Export of Bovine Blood Plasma/Serum for Manufacture into Pharmaceutical Products.



Authorized by
Derek Robinson
02 Dec 2013



Inspected by
Jennifer Ware
16 Jul 2015

