

New England Biolabs Certificate of Analysis

Product Name: BsrGI
Catalog Number: R0575L
Concentration: 10,000 U/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 µl.
Lot Number: 10030192
Expiration Date: 11/2020
Storage Temperature: -20°C
Storage Conditions: 50 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA
Specification Version: PS-R0575S/L v1.0

BsrGI Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0575LVIAL	BsrGI	10030194	Pass
B7202SVIAL	NEBuffer™ 2.1	10029698	Pass
B7024SVIAL	Gel Loading Dye, Purple (6X)	10021129	Pass


Assay Name/Specification	Lot # 10030192
<p>Endonuclease Activity (Nicking) A 50 µl reaction in NEBuffer 2.1 containing 1 µg of supercoiled PhiX174 DNA and a minimum of 50 Units of BsrGI incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.</p>	Pass
<p>Exonuclease Activity (Radioactivity Release) A 50 µl reaction in NEBuffer 2.1 containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 100 units of BsrGI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.</p>	Pass
<p>Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of Lambda DNA with BsrGI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with BsrGI.</p>	Pass
<p>Non-Specific DNase Activity (16 Hour) A 50 µl reaction in NEBuffer 2.1 containing 1 µg of Lambda DNA and a minimum of 100 Units of BsrGI incubated for 16 hours at 37°C results in a DNA pattern free of</p>	Pass

Assay Name/Specification	Lot # 10030192
detectable nuclease degradation as determined by agarose gel electrophoresis.	

This product has been tested and shown to be in compliance with all specifications.



Tony Spear-Alfonso
Production Scientist
08 Nov 2018



Michael Tonello
Packaging Quality Control Inspector
27 Nov 2018