

## New England Biolabs Certificate of Analysis

Product Name: *Oligo d(T)23VN*  
 Catalog Number: *S1327S*  
 Packaging Lot Number: *10058682*  
 Expiration Date: *07/2022*  
 Storage Temperature: *-20°C*  
 Storage Conditions: *Supplied as a lyophilized triethylammonium salt*  
 Specification Version: *PS-S1327S v1.0*

Oligo d(T)23VN Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
S1327SVIAL	Oligo d(T) <sub>23</sub> VN	10047681	Pass

Assay Name/Specification	Lot # 10058682
<b>Electrophoretic Pattern (Oligonucleotide)</b> The banding pattern of Oligo d(T)23 VN on a 15% TBE-Urea gel is evaluated against a control lot for expected size, sharpness, and relative intensity.	Pass

This product has been tested and shown to be in compliance with all specifications.



Mike Kuska  
Production Scientist  
05 Dec 2019

Michael Tonello  
Packaging Quality Control Inspector  
05 Dec 2019