

## New England Biolabs Certificate of Analysis

Product Name: *Oligo (dT)25-Cellulose*  
 Catalog Number: *S1408S*  
 Lot Number: *10040034*  
 Expiration Date: *04/2021*  
 Storage Temperature: *4°C*  
 Storage Conditions: *0.02% NaN<sub>3</sub>*  
 Specification Version: *PS-S1408S v1.0*

Oligo (dT)25-Cellulose Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
S1408SVIAL	Oligo (dT) <sub>25</sub> Cellulose Beads	10040035	Pass

Assay Name/Specification	Lot # 10040034
<b>Binding Capacity (Poly rA)</b> Oligo (dT)25Cellulose Beads (0.15 ml) were equilibrated and incubated with 0.5 ml of PolyA substrate solution for 5 minutes at 25°C, then washed and the PolyA eluted. Binding capacity was determined to be >400 OD per gram of cellulose.	Pass

This product has been tested and shown to be in compliance with all specifications.



John Buswell  
Production Scientist  
26 Apr 2019



Michael Tonello  
Packaging Quality Control Inspector  
26 Apr 2019