


## New England Biolabs Certificate of Analysis

Product Name: *Monarch DNA Cleanup Columns*  
 Catalog Number: *T1034L*  
 Concentration: *1 X Concentrate*  
 Lot Number: *10020894*  
 Expiration Date: *01/2021*  
 Storage Temperature: *25°C*  
 Specification Version: *PS-T1034L v1.0*

Assay Name/Specification	Lot # 10020894
<b>Functional Testing (DNA Recovery and Purity)</b> Twenty-four Monarch® DNA Cleanup Columns are tested with 1 µg of 1 Kb Ladder resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	<b>Pass</b>

This product has been tested and shown to be in compliance with all specifications.



Ana Egana  
Production Scientist  
09 Jan 2019



Heidi Church  
Packaging Quality Control Inspector  
09 Jan 2019