


New England Biolabs Certificate of Analysis

Product Name: *Monarch gDNA Removal Columns*
 Catalog Number: *T2017L*
 Lot Number: *10043831*
 Expiration Date: *01/2021*
 Storage Temperature: *25°C*
 Specification Version: *PS-T2017L v1.0*

Assay Name/Specification	Lot # 10043831
Functional Testing (DNA Recovery) Eight Monarch® gDNA Removal Columns are tested with 10 µg of Lambda/BstEII Marker resulting in ≥65% recovery in ≥80% of the samples.	Pass

This product has been tested and shown to be in compliance with all specifications.



Ana Egana
Production Scientist
05 Sep 2019



Heidi Church
Packaging Quality Control Inspector
05 Sep 2019