

New England Biolabs Certificate of Analysis

Product Name: Monarch® gDNA Removal Columns
 Catalog Number: T2017L
 Packaging Lot Number: 10166741
 Expiration Date: 12/2024
 Storage Temperature: 25°C
 Specification Version: PS-T2017L v2.0

Monarch® gDNA Removal Columns Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
T2018-1	Monarch® Collection Tubes II	10166717	Pass
T2017-1	Monarch® gDNA Removal Columns	10166715	Pass

Assay Name/Specification	Lot # 10166741
Functional Testing (DNA Recovery) Eight Monarch® gDNA Removal Columns are tested with 10 µg of Lambda/BstEII Marker resulting in ≥65% recovery in ≥80% of the samples.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Michael Burzyk
Production Scientist
11 Jan 2023



Nikki Samuelson
Packaging Quality Control Inspector
20 Apr 2023