


## New England Biolabs Certificate of Analysis

**Product Name:** Monarch® RNA Cleanup Columns (50 µg)  
**Catalog Number:** T2047L  
**Lot Number:** 10041696  
**Expiration Date:** 01/2021  
**Storage Temperature:** 25°C  
**Specification Version:** PS-T2047L v1.0

Assay Name/Specification	Lot # 10041696
<b>Functional Testing (RNA Recovery and Purity)</b> Twenty-four Monarch® RNA Cleanup Columns (T2047, 50 µg) are tested with 5 µg of rRNA resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	<b>Pass</b>

This product has been tested and shown to be in compliance with all specifications.



Ana Egana  
 Production Scientist  
 01 Oct 2019



Heidi Church  
 Packaging Quality Control Inspector  
 01 Oct 2019