


New England Biolabs Certificate of Analysis

Product Name: Monarch™ RNA (500ug) Column & Tubes
Catalog Number: T2057L
Lot Number: 0011806
Expiration Date: 00/0000
Storage Temperature: 25°C
Specification Version: PS-T2057L v1.0

Assay Name/Specification	Lot # 0011806
Functional Testing (RNA Recovery and Purity) Twenty-four Monarch® RNA Cleanup Columns (T2057, 500 µg) are tested with 50 µg of rRNA resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	Pass

This product has been tested and shown to be in compliance with all specifications.



Ana Egana
 Production Scientist
 20 Jul 2018



Heidi Church
 Packaging Quality Control Inspector
 20 Jul 2018