


New England Biolabs Certificate of Analysis

Product Name: *Monarch[®] RNA Cleanup Columns (500 µg)*
 Catalog Number: *T2057L*
 Packaging Lot Number: *10059921*
 Expiration Date: *08/2021*
 Storage Temperature: *25°C*
 Specification Version: *PS-T2057L v1.0*

| Assay Name/Specification | Lot # 10059921 |
|---|----------------|
| <p>Functional Testing (RNA Recovery and Purity) Twenty-four Monarch[®] RNA Cleanup Columns (T2057, 500 µg) are tested with 50 µg of rRNA resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.</p> | Pass |

This product has been tested and shown to be in compliance with all specifications.



Ana Egana
Production Scientist
26 Mar 2020

Darcie Spaulding
Packaging Quality Control Inspector
26 Mar 2020