


## New England Biolabs Certificate of Analysis

Product Name: *Monarch gDNA Elution Buffer*  
 Catalog Number: *T3016L*  
 Packaging Lot Number: *10068617*  
 Expiration Date: *05/2022*  
 Storage Temperature: *25°C*  
 Specification Version: *PS-T3016L v1.0*


Assay Name/Specification	Lot # 10068617
<b>Functional Testing (DNA Recovery and Purity)</b> Twenty-four Monarch® gDNA Purification Columns are tested with 5 µg of HindIII digested Lambda resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	<b>Pass</b>

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.



Ana Egana  
Production Scientist  
04 Jan 2021



Darcie Spaulding  
Packaging Quality Control Inspector  
04 Jan 2021