

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: NEBuffer™ 4

Catalog Number: B7004S

Concentration: 10 X Concentrate

Packaging Lot Number: 10225678
Expiration Date: 03/2026
Storage Temperature: -20°C

Specification Version: PS-B7004S v1.0

Composition (1X): 50 mM Potassium Acetate , 20 mM Tris-acetate, 10 mM Magnesium

Acetate, 1 mM DTT, (pH 7.9 @ 25°C)

NEBuffer™ 4 Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
B7004SVIAL	NEBuffer™ 4	10184703	Pass	

Assay Name/Specification	Lot # 10225678
Conductivity (buffers/solutions)	Pass
The conductivity of 10X NEBuffer 4 is between 36 and 54 mS at 25°C.	
Endonuclease Activity (Nicking, Buffer)	Pass
A 50 µl reaction in 1X NEBuffer 4 containing 1 µg of supercoiled PhiX174 DNA	
incubated for 4 hours at 37°C results in <10% conversion to the nicked form as	
determined by agarose gel electrophoresis.	
Functional Testing (Restriction Digest, BSA, Buffer)	Pass
A 50 µl reaction in 1X NEBuffer 4 plus 100 µg/ml Bovine Serum Albumin containing 1	
μg of Lambda dam- DNA and 1 unit of Clal incubated for 1 hour at 37°C results in	
complete digestion of the substrate DNA as determined by agarose gel	
electrophoresis.	
Functional Tacting (Pastwiction Digast BCA Buffer)	Pass
Functional Testing (Restriction Digest, BSA, Buffer)  A 50 µl reaction in 1X NEBuffer 4 plus 100 µg/ml Bovine Serum Albumin containing 1	Pass
µg of Lambda DNA and 1 unit of MscI incubated for 1 hour at 37°C results in complete	
digestion of the substrate DNA as determined by agarose gel electrophoresis.	
Non-Specific DNase Activity (16 hour, Buffer)	Pass
A 50 μl reaction in 1X NEBuffer 4 containing 1 μg of PhiX174-HaeIII DNA incubated	
for 16 hours at 37°C results in a DNA pattern free of detectable nuclease	



B7004S / Lot: 10225678

Page 1 of 2

The pH of 10X NEBuffer 4 is between pH 7.8 and 8.0 at 25°C.

**Pass** 

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Nancy Considine Production Scientist 12 May 2023

my Gondon

pH (buffers/solutions)

Michael Tonello

Packaging Quality Control Inspector

27 Dec 2023



B7004S / Lot: 10225678

Page 2 of 2