

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name:	NEBuffer TM 1.1
Catalog #:	<i>B7201S</i>
Concentration:	10X Concentrate
Lot #:	0121709
Assay Date:	09/2017
Expiration Date:	09/2020
Storage Temp:	-20°C
Composition (1X):	10 mM Bis-Tris Propane, 10 mM MgCl₂, 100 μg/ml BSA, (pH 7.0 @ 25°C)
Specification Version:	PS-B7201S v1.0
Effective Date:	31 Jan 2018

Assay Name/Specification (minimum release criteria)	Lot #0121709
Conductivity (buffers/solutions) - The conductivity of 10X NEBuffer 1.1 is between 23 and 26 mS at 25°C.	Pass
Endonuclease Activity (Nicking, Buffer) - A 50 μ l reaction in 1X NEBuffer 1.1 containing 1 μ g of supercoiled PhiX174 DNA incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Functional Testing (Restriction Digest, Buffer) - A 50 µl reaction in 1X NEBuffer 1.1 containing 1 µg of Lambda-HindIII DNA and 1 unit of SacI incubated for 1 hour at 37°C results in complete digestion of the substrate DNA as determined by agarose gel electrophoresis.	Pass
Functional Testing (Restriction Digest, Buffer) - A 50 µl reaction in 1X NEBuffer 1.1 containing 1 µg of pXba DNA and 1 unit of KpnI incubated for 1 hour at 37°C results in complete digestion of the substrate DNA as determined by agarose gel electrophoresis.	Pass
Non-Specific DNase Activity (16 hour, Buffer) - A 50 µl reaction in 1X NEBuffer 1.1 containing 1 µg of PhiX174-HaeIII DNA incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
pH (buffers/solutions) - The pH of 10X NEBuffer 1.1 is between pH 6.9 and 7.1 at 25°C.	Pass
RNase Activity (Buffer) - A 10 μ l reaction in 1X NEBuffer 1.1 containing 40 ng of a 300 base single-stranded RNA is incubated at 37°C. After incubation for 16 hours, >90% of the substrate RNA remains intact as determined by fluorescent detection.	Pass

* The BSA in this product has been granted an EDQM "Certificate of Suitability" from the European Directorate for the Quality of Medicines (# R1-CEP-2003-204-Rev00) and has been granted a USDA Certificate for Export of Bovine Blood Plasma/Serum for Manufacture into Pharmaceutical Products.

otimon

Authorized by Derek Robinson 31 Jan 2018



Allfor In

Inspected by Tony Spear-Alfonso 05 Oct 2017