

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: CutSmart® Buffer

Catalog Number: B7204S

Concentration: 10 X Concentrate

Lot Number: 10030519
Expiration Date: 10/2021
Storage Temperature: -20°C

Specification Version: PS-B7204S v1.0

Composition (1X): 50 mM Potassium Acetate , 20 mM Tris Acetate, 10 mM Magnesium

Acetate, 100 μg/ml BSA, (pH 7.9 @ 25°C)

| CutSmart® Buffer Component List | | | | |
|---------------------------------|-----------------------|------------|----------------------|--|
| NEB Part Number | Component Description | Lot Number | Individual QC Result | |
| B7204SVIAL | CutSmart® Buffer | 10021122 | Pass | |

| Assay Name/Specification | Lot # 10030519 |
|---|----------------|
| Non-Specific DNase Activity (16 hour, Buffer) | Pass |
| A 50 μl reaction in 1X CutSmart® Buffer containing 1 μg of PhiX174-HaeIII DNA | |
| incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis. | |
| degradation as determined by agarose gerelectrophoresis. | |
| pH (buffers/solutions) | Pass |
| The pH of 10X CutSmart® Buffer is between pH 7.8 and 8.0 at 25°C. | |
| | |
| Functional Testing (Restriction Digest, Buffer) | Pass |
| A 50 µl reaction in 1X CutSmart® Buffer containing 1 µg of Lambda DNA and 1 unit of | |
| MscI incubated for 1 hour at 37°C results in complete digestion of the substrate DNA as determined by agarose gel electrophoresis. | |
| as determined by against generate prioresis. | |
| Functional Testing (Restriction Digest, Buffer) | Pass |
| A 50 µl reaction in 1X CutSmart® Buffer containing 1 µg of Lambda dam- DNA and 1 | |
| unit of Clal incubated for 1 hour at 37°C results in complete digestion of the | |
| substrate DNA as determined by agarose gel electrophoresis. | |
| RNase Activity (Buffer) | Pass |
| A 10 µl reaction in 1X CutSmart® Buffer containing 40 ng of a 300 base | 1 433 |
| single-stranded RNA is incubated at 37°C. After incubation for 16 hours, >90% of the | |
| substrate RNA remains intact as determined by fluorescent detection. | |



B7204S / Lot: 10030519

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

Tony Spear-Alfonso Production Scientist

08 Nov 2018

Michael Tonello

Packaging Quality Control Inspector

21 Nov 2018



B7204S / Lot: 10030519

Page 2 of 2