

New England Biolabs Certificate of Analysis

Product Name: Diluent C
Catalog Number: B8003S
Concentration: 1 X Concentrate
Packaging Lot Number: 10170009
Expiration Date: 11/2025
Storage Temperature: -20°C
Specification Version: PS-B8003S v1.0
Composition (1X): 250 mM NaCl, 10 mM Tris-HCl, 1 mM DTT, 0.1 mM EDTA, 50 % Glycerol, 0.15 % Triton X-100, 200 µg/ml BSA, (pH 7.4 @ 25°C)

Diluent C Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
B8003SVIAL	Diluent C	10170008	Pass

Assay Name/Specification	Lot # 10170009
qPCR DNA Contamination (E. coli Genomic) A minimum of 1 µl of Diluent C is screened for the presence of E. coli genomic DNA using SYBR® Green qPCR with primers specific for the E. coli 16S rRNA locus. Results are quantified using a standard curve generated from purified E. coli genomic DNA. The measured level of E. coli genomic DNA contamination is ≤ 1 E. coli genome.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart® Buffer containing 1 µg of PhiX174-HaeIII DNA and a minimum of 10 µl of Diluent C incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Endonuclease Activity (Nicking) A 50 µl reaction in CutSmart® Buffer containing 1 µg of supercoiled PhiX174 DNA and a minimum of 10 µl of Diluent C incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
pH (buffers/solutions) The pH of 1X Diluent C is between pH 7.3 and 7.5 at 25°C.	Pass
RNase Activity (Extended Digestion) A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA	Pass

Assay Name/Specification	Lot # 10170009
and a minimum of 1 µl of Diluent C is incubated at 37°C. After incubation for 16 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Nancy Considine
Production Scientist
14 Nov 2022



Josh Hersey
Packaging Quality Control Inspector
15 Nov 2022