

## New England Biolabs Certificate of Analysis

Product Name: NEB® 10-beta Competent *E. coli* (High Efficiency)  
 Catalog Number: C3019H  
 Lot Number: 10020868  
 Expiration Date: 08/2019  
 Storage Temperature: -80°C  
 Specification Version: PS-C3019H/I v1.0


| NEB® 10-beta Competent <i>E. coli</i> (High Efficiency) Component List |   |            |                      |
|--|---|------------|----------------------|
| NEB Part Number  | Component Description                                   | Lot Number | Individual QC Result |
| N3041AVIAL   | pUC19 Vector  | 10017063   | Pass                 |
| C3019HVIAL   | NEB® 10-beta Competent <i>E. coli</i> (High Efficiency) | 10016111   | Pass                 |
| B9035SVIAL   | NEB® 10-beta/Stable Outgrowth Medium                    | 0341806    | Pass                 |

| Assay Name/Specification   | Lot # 10020868 |
|--|----------------|
| <p><b>Phage Resistance (<math>\phi</math> 80)</b><br/>           15 <math>\mu</math>l of untransformed NEB® 10-beta Competent <i>E. coli</i> (High Efficiency) streaked onto a Rich Broth plate does not support plaque formation by phage <math>\phi</math> 80 after incubation for 16 hours at 37°C.</p>   | Pass           |
| <p><b>Transformation Efficiency</b><br/>           50 <math>\mu</math>l of NEB® 10-beta Competent <i>E. coli</i> (High Efficiency) cells were transformed with 100 pg of pUC19 DNA using the transformation protocol provided. Incubation overnight on LB-Ampicillin plates at 37°C resulted in <math>&gt;1 \times 10^9</math> cfu/<math>\mu</math>g of DNA.</p> | Pass           |
| <p><b>Antibiotic Resistance (Streptomycin)</b><br/>           15 <math>\mu</math>l of untransformed NEB® 10-beta Competent <i>E. coli</i> (High Efficiency) streaked onto a Rich Broth plate containing Streptomycin will form colonies after incubation for 16 hours at 37°C.</p>   | Pass           |
| <p><b>Antibiotic Sensitivity (Ampicillin)</b><br/>           15 <math>\mu</math>l of untransformed NEB® 10-beta Competent <i>E. coli</i> (High Efficiency) streaked onto a Rich Broth plate containing Ampicillin will not form colonies after incubation for 16 hours at 37°C.</p>  | Pass           |
| <p><b>Antibiotic Sensitivity (Chloramphenicol)</b><br/>           15 <math>\mu</math>l of untransformed NEB® 10-beta Competent <i>E. coli</i> (High Efficiency) streaked onto a Rich Broth plate containing Chloramphenicol will not form colonies after</p>   | Pass           |

| Assay Name/Specification   | Lot # 10020868 |
|--|----------------|
| <p>incubation for 16 hours at 37°C.</p> <p><b>Antibiotic Sensitivity (Kanamycin)</b><br/>15 µl of untransformed NEB® 10-beta Competent E. coli (High Efficiency) streaked onto a Rich Broth plate containing Kanamycin will not form colonies after incubation for 16 hours at 37°C.</p> | <b>Pass</b>    |
| <p><b>Antibiotic Sensitivity (Nitrofurantoin)</b><br/>15 µl of untransformed NEB® 10-beta Competent E. coli (High Efficiency) streaked onto a Rich Broth plate containing Nitrofurantoin will not form colonies after incubation for 16 hours at 37°C.</p>                               | <b>Pass</b>    |
| <p><b>Antibiotic Sensitivity (Spectinomycin)</b><br/>15 µl of untransformed NEB® 10-beta Competent E. coli (High Efficiency) streaked onto a Rich Broth plate containing Spectinomycin will not form colonies after incubation for 16 hours at 37°C.</p>                                 | <b>Pass</b>    |
| <p><b>Antibiotic Sensitivity (Tetracycline)</b><br/>15 µl of untransformed NEB® 10-beta Competent E. coli (High Efficiency) streaked onto a Rich Broth plate containing Tetracycline will not form colonies after incubation for 16 hours at 37°C.</p>                                   | <b>Pass</b>    |
| <p><b>Blue-White Screening (α-complementation, Competent Cells)</b><br/>NEB® 10-beta Competent E. coli (High Efficiency) were shown to be suitable for blue/white screening by α-complementation of the β-galactosidase gene using pUC19.</p>  | <b>Pass</b>    |

This product has been tested and shown to be in compliance with all specifications.

Chris Blanchette  
Production Scientist  
27 Jul 2018



Nick Privitera  
Packaging Quality Control Inspector  
30 Aug 2018