

*be* INSPIRED *drive* DISCOVERY *stay* GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name:	HiScribe™ SP6 RNA Synthesis Kit
Catalog Number:	E2070S
Packaging Lot Number:	10088667
Expiration Date:	02/2022
Storage Temperature:	-20°C
Specification Version:	PS-E2070S v1.0

HiScribe™ SP6 RNA Synthesis Kit Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
N2074AVIAL	CTP (Tris)	10081698	Pass	
N2073AVIAL	UTP (Tris)	10081697	Pass	
N2072AVIAL	GTP (Tris)	10081696	Pass	
N2071AVIAL	ATP (Tris)	10081695	Pass	
N2070AVIAL	SP6 Control Template	10071300	Pass	
M2070AVIAL	SP6 RNA Polymerase Mix	10081694	Pass	
M0303AVIAL	DNase I (RNase-free)	10081693	Pass	
B2070AVIAL	10X SP6 Reaction Buffer	10081699	Pass	
B2051AVIAL	LiCI Solution	10066827	Pass	

Assay Name/Specification	Lot # 10088667
Functional Testing (RNA Synthesis, Yield)	Pass
A 25 µl reaction in SP6 Reaction Buffer in the presence of 20 mM NTPs containing 1	
μg SP6 Control Template and 2.5 μl of the SP6 RNA Polymerase Mix incubated for 2	
hours at 37°C results in $\ge$ 80 µg of a 1.75 kb RNA as determined by OD 260 and	
agarose gel electrophoresis.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.





*be* INSPIRED *drive* DISCOVERY *stay* GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

fue ongeran

Dongxian Yue Production Scientist 21 Oct 2020

Michael	Jon	llo
---------	-----	-----

Michael Tonello Packaging Quality Control Inspector 02 Nov 2020

