

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: LunaScript RT SuperMix Kit

Catalog Number: E3010L
Packaging Lot Number: 10138628
Expiration Date: 02/2024
Storage Temperature: -20°C

Specification Version: PS-E3010S/L/G v1.0

LunaScript RT SuperMix Kit Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
M3011LVIAL	No-RT Control Mix	10135651	Pass
M3010LVIAL	LunaScript® RT SuperMix	10135650	Pass
B1502AVIAL	Nuclease-free Water	10135544	Pass

Assay Name/Specification	Lot # 10138628 Pass
Functional Testing (Two-Step RT-qPCR) The LunaScript™ RT SuperMix Kit is functionally tested in two-step RT-qPCR with human RNA template, resulting in a standard curve with a calculated qPCR efficiency of 90-110%, and a dynamic range of 7 orders of magnitude.	
* Individual Product Component Note Standard Quality Control Tests are performed for each component included in LunaScript™ RT SuperMix Kit and meet the designated specifications.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



E3010L / Lot: 10138628

Page 1 of 2

Christie Vazquez
Production Scientist

09 Feb 2022

Josh Hersey

Packaging Quality Control Inspector

09 Feb 2022