

New England Biolabs Certificate of Analysis

Product Name: NEBNext UltraExpress® RNA Library Prep Kit
Catalog Number: E3330L
Packaging Lot Number: 10257191
Expiration Date: 01/2026
Storage Temperature: -20°C
Specification Version: PS-E3330L v1.0

NEBNext UltraExpress® RNA Library Prep Kit Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7762AAVIAL	NEBNext® Adaptor Dilution Buffer	10248909	Pass
E3342AAVIAL	Nuclease-free Water	10248911	Pass
E3341AAVIAL	0.1X TE	10248910	Pass
E3339AAVIAL	NEBNext® Bead Reconstitution Buffer	10248912	Pass
E3338AAVIAL	NEBNext® MSTC™ High Yield Master Mix	10248908	Pass
E3337AAVIAL	NEBNext UltraExpress® USER Enzyme	10248907	Pass
E3336AAVIAL	NEBNext UltraExpress® Ligation Master Mix	10248906	Pass
E3335AAVIAL	NEBNext UltraExpress® End Prep Reaction Buffer	10248905	Pass
E3334AAVIAL	NEBNext UltraExpress® End Prep Enzyme Mix	10248904	Pass
E3333AAVIAL	NEBNext UltraExpress® Second Strand Master Mix	10248903	Pass
E3332AAVIAL	NEBNext UltraExpress® Strand Specificity Reagent	10248902	Pass
E3331AAVIAL	NEBNext UltraExpress® First Strand Enzyme Mix	10248901	Pass
E3329AAVIAL	NEBNext UltraExpress® RNA Fragmentation Mix	10248900	Pass

Assay Name/Specification	Lot # 10257191
* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext UltraExpress™ RNA Library Prep Kit and meet the designated specifications.	Pass
Functional Testing (Library Construction, RNA) Each set of reagents is functionally validated and compared to a previous lot through construction of libraries made from commercially available RNA, using the	Pass

Assay Name/Specification	Lot # 10257191
kit's minimum and maximum input requirements. Libraries made from previous and current lots for both input RNA amounts are sequenced together on the same Illumina® flow cell and compared across various metrics including library yield, individual transcript abundance correlations (low vs. high input, old lot vs. new lot), 5'-3' transcript coverage, and fraction of reads mapping to a reference.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Christine Sumner
Production Scientist
04 Sep 2024



Josh Hersey
Packaging Quality Control Inspector
04 Sep 2024