

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: OneTaq® One-Step RT-PCR Kit

Catalog Number: E5315S
Packaging Lot Number: 10203984
Expiration Date: 07/2024
Storage Temperature: -20°C

Specification Version: PS-E5315S v1.0

OneTaq® One-Step RT-PCR Kit Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
M1205AVIAL	Quick Load® OneTaq® One-Step Reaction Mix	10200595	Pass	
M1204AVIAL	OneTaq® One-Step Reaction Mix	10200594	Pass	
M1203AVIAL	OneTaq® One-Step Enzyme Mix	10200593	Pass	
B1502AVIAL	Nuclease-free Water	10185496	Pass	

Assay Name/Specification	Lot # 10203984
* Individual Product Component Note Standard Quality Control Tests are performed for each component included in OneTaq® One-Step RT-PCR Kit and meet the designated specifications.	Pass
Functional Testing (RT-PCR, Length) A 50 μl reaction in OneTaq® One-Step Reaction Mix or Quick-Load® OneTaq® One-Step Reaction Mix containing 100 ng of total RNA, 1X OneTaq® One-Step Enzyme Mix, and 0.4 μM primers for 40 cycles of PCR amplification results in the expected 9.3 kb product.	Pass
Functional Testing (RT-PCR, Multiplexing) A 50 μl reaction in OneTaq® One-Step Reaction Mix or Quick-Load® OneTaq® One-Step Reaction Mix containing 10 ng of total RNA, 1X OneTaq® One-Step Enzyme Mix, and 0.1 μM primers for 40 cycles of PCR amplification results in the expected three products.	Pass
Functional Testing (RT-PCR, Sensitivity) A 50 μl reaction in OneTaq® One-Step Reaction Mix or Quick-Load® OneTaq® One-Step Reaction Mix containing 10 pg of total RNA, 1X OneTaq® One-Step Enzyme Mix, and 0.4 μM primers for 40 cycles of PCR amplification results in the expected 566 bp product.	Pass

This product has been tested and shown to be in compliance with all specifications.



E5315S / Lot: 10203984 Page 1 of 2



Lea Antonpoulos Production Scientist

11 Aug 2023

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Michael Tonello

Packaging Quality Control Inspector

24 Aug 2023



E5315S / Lot: 10203984

Page 2 of 2