

New England Biolabs Certificate of Analysis


Product Name: *NEBuilder[®] HiFi DNA Assembly Cloning Kit*
 Catalog Number: *E5520S*
 Lot Number: *10014838*
 Expiration Date: *06/2019*
 Storage Temperature: *Multi-temperature**
 Specification Version: *PS-E5520S v1.0*

* This product contains components with different storage temperature requirements. Please reference the applicable product specification document(s) on the Quality and Safety tab located on the product page of www.neb.com.

NEBuilder [®] HiFi DNA Assembly Cloning Kit Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
N3041AVIAL	pUC19 Vector	10010689	Pass
N2611AVIAL	Gibson Assembly [®] Positive Control	0171802	Pass
M5520AVIAL	NEBuilder [®] High-Fidelity Master Mix	10009010	Pass
C2987HVIAL	NEB [®] 5-alpha Competent E. coli (High Efficiency)	10009119	Pass
B9020SVIAL	SOC Outgrowth Medium	5671804	Pass

Assay Name/Specification	Lot # 10014838
<p>* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBuilder[®] HiFi DNA Assembly Cloning Kit and meet the designated specifications.</p> <p>Functional Testing (NEBuilder[®] HiFi DNA Assembly) 10 µl of 2X NEBuilder[®] HiFi DNA Assembly Cloning Kit was incubated with 0.05 pmol each of 6 DNA fragments (4 fragments of 1,000 bp, one fragment of 1,152 bp with 80 bp overlap, and a vector of 3,373 bp with a 20 bp overlap) in a final volume of 20 µl at 50°C for 60 minutes. NEB[®] 5-alpha Competent E. coli (High Efficiency) were transformed with 2 µl of the assembled products. Successfully assembled fragments produce an intact lacZ gene in the pACYC184 vector, and yield blue colonies on an IPTG/Xgal/Chloramphenicol plate when incubated overnight at 37°C after transformation. Greater than 100 blue colonies were observed when 1/10 of the outgrowth (500 µl) was spread on a plate.</p>	<p>Pass</p> <p>Pass</p>

This product has been tested and shown to be in compliance with all specifications.



Ana Egana
Production Scientist
11 Jul 2018



Nick Privitera
Packaging Quality Control Inspector
11 Jul 2018