

be INSPIRED drive DISCOVERY stay GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

| NEBNext® rRNA Depletion Kit (Human/Mouse/Rat) |
|---|
| E6310L |
| 10065070 |
| 06/2021 |
| -20°C |
| PS-E6310S/L/X/B v1.0 |
| |

| NEBNext® rRNA Depletion Kit (Human/Mouse/Rat) Component List | | | |
|--|-------------------------------------|------------|----------------------|
| NEB Part Number | Component Description | Lot Number | Individual QC Result |
| E6318-3VIAL | NEBNext® RNase H | 10043015 | Pass |
| E6317-3VIAL | Nuclease-free Water | 10043014 | Pass |
| E6316-3VIAL | DNase I (RNase-free) | 10043013 | Pass |
| E6315-3VIAL | DNase I Reaction Buffer | 10043012 | Pass |
| E6314-3VIAL | NEBNext® Probe Hybridization Buffer | 10043011 | Pass |
| E6313-3VIAL | NEBNext® rRNA Depletion Solution | 10043010 | Pass |
| E6312-3VIAL | RNase H Reaction Buffer | 10043009 | Pass |

| Assay Name/Specification | Lot # 10065070 |
|--|----------------|
| * Individual Product Component Note | Pass |
| Standard Quality Control Tests are performed for each component included in NEBNext® | |
| rRNA Depletion Kit (Human/Mouse/Rat) and meet the designated specifications. | |
| Functional Testing (rRNA Depletion) | Pass |
| The NEBNext rRNA Depletion Kit is functionally validated using commercially | |
| available human RNA (e.g. UHRR). After treatment with the kit, a high throughput | |
| sequencing library is made and sequenced. Reads from this library are classified as | |
| being derived from ribosomal RNA based on sequence homology. This method produces | |
| libraries with less than 10% rRNA. | |

This product has been tested and shown to be in compliance with all specifications.





be INSPIRED drive DISCOVERY stay GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

notin ~

Christine Sumner Production Scientist 24 Jan 2020

Michae m. l

Michael Tonello Packaging Quality Control Inspector 24 Jan 2020

