

New England Biolabs Certificate of Analysis

Product Name: NEBNext Multiplex Oligos for Illumina (96 Unique Dual Index Primer Pairs)
Catalog Number: E6440S
Packaging Lot Number: 10071303
Expiration Date: 04/2022
Storage Temperature: -20°C
Specification Version: PS-E6440S/L v1.0

| NEBNext Multiplex Oligos for Illumina (96 Unique Dual Index Primer Pairs) Component List | | | |
|--|--|------------|----------------------|
| NEB Part Number | Component Description | Lot Number | Individual QC Result |
| E6612AVIAL | NEBNext® Adaptor for Illumina® | 10071306 | Pass |
| E6610AVIAL | USER Enzyme | 10071304 | Pass |
| E6441AVIAL | NEBNext Unique Dual Index Primer Pairs | 10071305 | Pass |

| Assay Name/Specification | Lot # 10071303 |
|--|----------------|
| <p>* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext® Multiplex Oligos for Illumina® (96 Unique Dual Index Primer Pairs) and meet the designated specifications.</p> | Pass |
| <p>Functional Testing (Library Construction, Oligo) Each of the components are functionally validated through construction of libraries made from commercially available genomic DNA. Libraries are made from each unique barcoded oligonucleotide and are sequenced together on the same Illumina flow cell and compared across various metrics including library yield, fraction of reads aligning to the reference, GC bias, and insert size.</p> | Pass |

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Christine Sumner

Christine Sumner
Production Scientist
20 Apr 2020

Michael Tonello

Michael Tonello
Packaging Quality Control Inspector
14 Sep 2020