

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: ProtoScript® II First Strand cDNA Synthesis Kit

Catalog Number: E6560L
Packaging Lot Number: 10219573
Expiration Date: 03/2025
Storage Temperature: -20°C

Specification Version: PS-E6560S/L v1.0

| ProtoScript® II First Strand cDNA Synthesis Kit Component List |   |            |                      |
|--|---|------------|----------------------|
| <b>NEB Part Number</b>   | Component Description                           | Lot Number | Individual QC Result |
| E6560S   | ProtoScript® II First Strand cDNA Synthesis Kit | 10219572   | Pass                 |

| Assay Name/Specification   | Lot # 10219573 |
|--|----------------|
| * Individual Product Component Note  | Pass           |
| Standard Quality Control Tests are performed for each component included in          |                |
| ProtoScript® II First Strand cDNA Synthesis Kit and meet the designated              |                |
| specifications.  |                |
| Functional Testing (RT-PCR, Length)  | Pass           |
| The performance of ProtoScript® II Reaction Mix with 1X ProtoScript® II Enzyme Mix   |                |
| is tested in a 20 μl RT reaction with 0.5 μg Human total RNA. The length of the      |                |
| product is verified by amplification of fibrillin using 0.4 µl of the RT reaction    |                |
| and 40 cycles of PCR amplification resulting in the expected 9.2 kb product.         |                |
| Functional Testing (RT-PCR, Sensitivity)   | Pass           |
| The performance of ProtoScript® II Reaction Mix with 1X ProtoScript® II Enzyme Mix   |                |
| is tested in a 20 µl RT reaction with 20 pg Human total RNA. The sensitivity of the  |                |
| product is verified by amplification of GAPDH using 0.4 µl of the RT reaction and 35 |                |
| cycles of PCR amplification resulting in the expected product.                       |                |

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



E6560L / Lot: 10219573

Page 1 of 2

Christie Vazquez **Production Scientist** 

21 Dec 2023

Josh Hersey Packaging Quality Control Inspector

21 Dec 2023

E6560L / Lot: 10219573

Page 2 of 2