

be INSPIRED drive DISCOVERY stay GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name:	PURExpress® (aa, tRNA) Kit
Catalog Number:	E6840S
Packaging Lot Number:	10118297
Expiration Date:	08/2023
Storage Temperature:	-80°C
Specification Version:	PS-E6840S v2.0

PURExpress® (aa, tRNA) Kit Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
P0760AVIAL	PURExpress <sup>®</sup> Solution B	10118302	Pass
N6843AVIAL	Amino Acid Mix	10118300	Pass
N6842AVIAL	PURExpress® tRNA Mix	10118299	Pass
N0424AVIAL	PURExpress Control DHFR Plasmid	10103267	Pass
B6841AVIAL	PURExpress® delta aa, tRNA	10118298	Pass

Assay Name/Specification	Lot # 10118297
Functional Testing (Cell Free Protein Synthesis Assay) (galactosidase) A 25 $\mu$ l reaction in the presence of 250 ng $\beta$ -galactosidase template DNA and 20 units RNase Inhibitor containing the components of the PURExpress® $\Delta$ (aa, tRNA) Kit incubated for 2 hours at 37°C results in the expected 116 kDa product as determined by SDS-PAGE with Coomassie Blue detection.	Pass
Functional Testing (Cell Free Protein Synthesis Assay) (Vent DNA Polymerase A 25 $\mu$ I reaction in the presence of 250 ng Vent DNA Polymerase template DNA and 20 units RNase Inhibitor containing the components of the PURExpress® $\Delta$ (aa, tRNA) Kit incubated for 2 hours at 37°C results in the expected 89 kDa product as determined by SDS-PAGE with Coomassie Blue detection.	Pass
* Individual Product Component Note Standard Quality Control Tests are performed for each component included in PURExpress® $\Delta$ (aa, tRNA) Kit and meet the designated specifications.	Pass
Functional Testing (Cell Free Protein Synthesis Assay) (DHFR) A 25 $\mu$ l reaction in the presence of 250 ng PURExpress® Control DHFR Plasmid and 20 units RNase Inhibitor containing the components of the PURExpress® $\Delta$ (aa, tRNA) Kit incubated for 2 hours at 37°C results in the expected 20 kDa product as determined by SDS-PAGE with Coomassie Blue detection.	Pass





be INSPIRED drive DISCOVERY stay GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Orinna Tuckey

Cory Tuckey Production Scientist 25 Aug 2021

Michae 011.

Michael Tonello Packaging Quality Control Inspector 25 Aug 2021

