

New England Biolabs Certificate of Analysis

Product Name: NEBNext® Companion Module for Oxford Nanopore Technologies® Ligation Sequencing
Catalog Number: E7180L
Packaging Lot Number: 10231267
Expiration Date: 03/2025
Storage Temperature: -20°C
Specification Version: PS-E7180S/L v2.0

NEBNext® Companion Module for Oxford Nanopore Technologies® Ligation Sequencing Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7185AAVIAL	Quick T4 DNA Ligase	10231272	Pass
E7184AAVIAL	NEBNext® Ultra II End Prep Enzyme Mix	10231271	Pass
E7183AAVIAL	NEBNext® Ultra II End Prep Reaction Buffer	10231270	Pass
E7182AAVIAL	NEBNext® FFPE DNA Repair Mix	10231269	Pass
E7181AAVIAL	NEBNext® FFPE DNA Repair Buffer	10231268	Pass

Assay Name/Specification	Lot # 10231267
<p>* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext® Companion Module for Oxford Nanopore Technologies® Ligation Sequencing and meet the designated specifications.</p>	Pass
<p>Functional Testing (Library Construction) Each set of reagents is functionally validated through the construction of libraries made using commercially available DNA and the kit's input requirement. Libraries are sequenced on individual Oxford Nanopore® flow cells and their quality assessed across a number of metrics including library yield, sequence yield, pore occupancy, and read length.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Christine Sumner

Christine Sumner
Production Scientist
12 Apr 2024

Michael Tonello

Michael Tonello
Packaging Quality Control Inspector
16 Apr 2024