

New England Biolabs Certificate of Analysis

Product Name: NEBNext® Ultra™ RNA Library Prep Kit for Illumina®
 Catalog Number: E7530S
 Lot Number: 10041072
 Expiration Date: 02/2020
 Storage Temperature: -20°C
 Specification Version: PS-E7530S/L v1.0

NEBNext® Ultra™ RNA Library Prep Kit for Illumina® Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7431AVIAL	Nuclease-free Water	10019117	Pass
E7427AVIAL	NEBNext® Second Strand Synthesis Reaction Buffer, 10X	10019118	Pass
E7425AVIAL	NEBNext® Second Strand Synthesis Enzyme Mix	10019113	Pass
E7424AVIAL	RNase Inhibitor, Murine	10019112	Pass
E7423AVIAL	ProtoScript® II Reverse Transcriptase	10019119	Pass
E7422AVIAL	Random Primers	10019111	Pass
E7421AVIAL	NEBNext® First Strand Synthesis Reaction Buffer	10019110	Pass
E7373AVIAL	Blunt/TA Ligase Master Mix	10019116	Pass
E7372AVIAL	NEBNext® End Repair Reaction Buffer	10019115	Pass
E7371AVIAL	NEBNext® End Prep Enzyme Mix	10019114	Pass
E6625AVIAL	NEBNext® Q5® Hot Start HiFi PCR Master Mix	10019120	Pass

Assay Name/Specification	Lot # 10041072
Functional Testing (Library Construction, RNA) Each set of reagents is functionally validated and compared to a previous lot through construction of libraries made from commercially available RNA, using the kit's minimum and maximum input requirements. Libraries made from previous and current lots for both input RNA amounts are sequenced together on the same Illumina® flow cell and compared across various metrics including library yield, individual transcript abundance correlations (low vs. high input, old lot vs. new lot), 5'-3' transcript coverage, and fraction of reads mapping to a reference.	Pass
* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext® Ultra™ RNA Library Prep Kit for Illumina® and meet the designated specifications.	Pass

This product has been tested and shown to be in compliance with all specifications.





Christine Sumner
Production Scientist
21 Mar 2019



Josh Hersey
Packaging Quality Control Inspector
21 Mar 2019