

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: NEBNext® Ultra™ II Directional RNA Second Strand Synthesis Module

Catalog Number: E7550L
Packaging Lot Number: 10157029
Expiration Date: 07/2023
Storage Temperature: -20°C

Specification Version: PS-E7550S/L v1.0

NEBNext® Ultra™ II Directional RNA Second Strand Synthesis Module Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7426AAVIAL	NEBNext® Second Strand Synthesis Reaction Buffer (dUTP Mix)	10146426	Pass
E7425AAVIAL	NEBNext® Second Strand Synthesis Enzyme Mix	10146425	Pass

Assay Name/Specification	Lot # 10157029
* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext® Ultra™ II Directional RNA Second Strand Synthesis Module and meet the designated	Pass
specifications.	
Functional Testing (Library Construction, RNA)	Pass
Each set of reagents is functionally validated and compared to the previous lot	
through construction of libraries made from commercially available RNA, using the	
kit's minimum and maximum input requirements. Libraries made from the previous and	
current lots for both input RNA amounts are sequenced together on the same Illumina	
flow cell and compared across various metrics including library yield, individual	
transcript abundance correlations (low vs. high input, old lot vs. new lot), 5'-3'	
transcript coverage, and fraction of reads mapping to a reference.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



E7550L / Lot: 10157029

Page 1 of 2



240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

Christin Summ

Christine Sumner Production Scientist 24 Jun 2022 Michael Tonello

Packaging Quality Control Inspector

24 Jun 2022



E7550L / Lot: 10157029

Page 2 of 2