

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: NEBNext® Ultra™ II Directional RNA Second Strand Synthesis Module

Catalog Number: E7550S
Packaging Lot Number: 10099480
Expiration Date: 10/2021
Storage Temperature: -20°C

Specification Version: PS-E7550S/L v1.0

NEBNext® Ultra™ II Directional RNA Second Strand Synthesis Module Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
	NEBNext® Second Strand Synthesis Reaction Buffer (dUTP Mix)	10071773	Pass	
E7425AVIAL	NEBNext® Second Strand Synthesis Enzyme Mix	10071772	Pass	

Assay Name/Specification	Lot # 10099480
Functional Testing (Library Construction, RNA)	Pass
Each set of reagents is functionally validated and compared to the previous lot	
through construction of libraries made from commercially available RNA, using the kit's minimum and maximum input requirements. Libraries made from the previous and	
current lots for both input RNA amounts are sequenced together on the same Illumina	
flow cell and compared across various metrics including library yield, individual	
transcript abundance correlations (low vs. high input, old lot vs. new lot), 5'-3'	
transcript coverage, and fraction of reads mapping to a reference.	
* Individual Product Component Note	Pass
Standard Quality Control Tests are performed for each component included in NEBNext®	
Ultra™ II Directional RNA Second Strand Synthesis Module and meet the designated	
specifications.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



E7550S / Lot: 10099480

Page 1 of 2



240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

Christin Summ

Christine Sumner Production Scientist 03 Feb 2021 Michael Tonello

Packaging Quality Control Inspector

03 Feb 2021



E7550S / Lot: 10099480

Page 2 of 2