

## New England Biolabs Certificate of Analysis

**Product Name:** NEBNext® Multiplex Oligos for Illumina® (Dual Index Primers Set 1)  
**Catalog Number:** E7600S  
**Packaging Lot Number:** 10236787  
**Expiration Date:** 02/2026  
**Storage Temperature:** -20°C  
**Specification Version:** PS-E7600S v1.0

NEBNext® Multiplex Oligos for Illumina® (Dual Index Primers Set 1) Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7622AVIAL	NEBNext® i712 Primer	10228640	Pass
E7621AVIAL	NEBNext® i711 Primer	10228639	Pass
E7620AVIAL	NEBNext® i710 Primer	10228638	Pass
E7619AVIAL	NEBNext® i709 Primer	10228637	Pass
E7618AVIAL	NEBNext® i708 Primer	10228636	Pass
E7617AVIAL	NEBNext® i707 Primer	10228635	Pass
E7616AVIAL	NEBNext® i706 Primer	10228634	Pass
E7615AVIAL	NEBNext® i705 Primer	10228633	Pass
E7614AVIAL	NEBNext® i704 Primer	10228632	Pass
E7613AVIAL	NEBNext® i703 Primer	10228631	Pass
E7612AVIAL	NEBNext® i702 Primer	10228630	Pass
E7611AVIAL	NEBNext® i701 Primer	10228629	Pass
E7610AVIAL	NEBNext® i508 Primer	10228628	Pass
E7609AVIAL	NEBNext® i507 Primer	10228627	Pass
E7608AVIAL	NEBNext® i506 Primer	10228626	Pass
E7607AVIAL	NEBNext® i505 Primer	10228625	Pass
E7606AVIAL	NEBNext® i504 Primer	10228624	Pass
E7605AVIAL	NEBNext® i503 Primer	10228623	Pass
E7604AVIAL	NEBNext® i502 Primer	10228622	Pass
E7603AVIAL	NEBNext® i501 Primer	10228621	Pass
E7602AVIAL	USER Enzyme	10228620	Pass
E7601AVIAL	NEBNext® Adaptor for Illumina®	10228619	Pass

<b>Assay Name/Specification</b>	<b>Lot # 10236787</b>
<b>* Individual Product Component Note</b>	<b>Pass</b>

Assay Name/Specification	Lot # 10236787
<p>Standard Quality Control Tests are performed for each component included in NEBNext® Multiplex Oligos for Illumina® (Dual Index Primers Set 1) and meet the designated specifications.</p> <p><b>Functional Testing (Library Construction, Oligo)</b> Each of the components are functionally validated through construction of libraries made from commercially available genomic DNA. Libraries are made from each unique barcoded oligonucleotide and are sequenced together on the same Illumina® flow cell and compared across various metrics including library yield, fraction of reads aligning to the reference, GC bias, and insert size.</p>	<p><b>Pass</b></p>

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.



Christine Sumner  
Production Scientist  
19 Mar 2024



Michael Tonello  
Packaging Quality Control Inspector  
19 Mar 2024