

## New England Biolabs Certificate of Analysis

Product Name: NEBNext<sup>®</sup> Globin & rRNA Depletion Kit (Human/Mouse/Rat)  
 Catalog Number: E7750S  
 Lot Number: 10049716  
 Expiration Date: 05/2021  
 Storage Temperature: -20°C  
 Specification Version: PS-E7750S/L/X v1.0

NEBNext <sup>®</sup> Globin & rRNA Depletion Kit (Human/Mouse/Rat) Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7753-2VIAL	NEBNext <sup>®</sup> DNase I	10042855	Pass
E7752-2VIAL	NEBNext <sup>®</sup> Thermostable RNase H	10042856	Pass
E7751-2VIAL	NEBNext <sup>®</sup> Globin & rRNA Depletion Solution	10042857	Pass
E6317-2VIAL	Nuclease-free Water	10042860	Pass
E6315-2VIAL	DNase I Reaction Buffer	10042859	Pass
E6314-2VIAL	NEBNext <sup>®</sup> Probe Hybridization Buffer	10042861	Pass
E6312-2VIAL	RNase H Reaction Buffer	10042858	Pass

Assay Name/Specification	Lot # 10049716
<p><b>Functional Testing (Globin and rRNA Depletion)</b></p> <p>The NEBNext<sup>®</sup> Globin &amp; rRNA Depletion Kit (Human/Mouse/Rat) is functionally validated using commercially available human RNA extracted from blood. After treatment with the kit, a high throughput sequencing library is made and sequenced. Reads from this library are classified as being derived from ribosomal RNA based on sequence homology. This method produces libraries with less than 10% rRNA and less than 5% globin RNA.</p>	Pass
<p><b>* Individual Product Component Note</b></p> <p>Standard Quality Control Tests are performed for each component included in NEBNext<sup>®</sup> Globin &amp; rRNA Depletion Kit (Human/Mouse/Rat) and meet the designated specifications.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

*Christine Sumner*

---

Christine Sumner  
Production Scientist  
25 Jul 2019

*Michael Tonello*

---

Michael Tonello  
Packaging Quality Control Inspector  
25 Jul 2019